





& Case Photos Provided by Istvan Urban DMD, MD, PhD



1. Labial view of an atrophic posterior mandibular area.



2. A 1:1 mixture of autogenous and xenogenic bone graft is placed on the ridge. Cortical bone was perforated, and an RPM Reinforced PTFE Mesh was secured on the lingual side before applying bone graft.



3. An RPM is secured over the graft with titanium pins and screws.



4. After 9 months of healing the augmented site is exposed, and the RPM will be removed.



5. & 6. Labial and occlusal views of the regenerated bone after 9 months of healing.





7. & 8. Labial and occlusal views of two implants placed into regenerated bone.





Versatile Rectangular Shapes

These configurations can be trimmed to fit a variety of defects



PS 20 mm x 25 mm



25 mm x 30 mm



XL 30 mm x 40 mm



XLK 30 mm x 40 mm



XLKM 30 mm x 40 mm



K240 mm x 50 mm

Shapes with Fixation Points

These configurations are designed with fixation points outside of the defect area



BL 17 mm x 25 mm



PST 25 mm x 36 mm



PLT 30 mm x 41 mm

Interproximal Shapes

These configurations are designed to fit between existing teeth



24 mm x 38 mm



AIGWI 24 mm x 38 mm



38 mm x 38 mm



PTCM 38 mm x 38 mm



PD 38 mm x 38 mm



PDMR 38 mm x 38 mm



PDML 38 mm x 38 mm